

AlGaN-based pH-Sensors

Impedance Characterisation, Optimisation and
Application for Foetal Blood Sampling

Benedikt Lübbers



Universitätsverlag Ilmenau
2013

Impressum

Bibliografische Information der Deutschen Nationalbibliothek

Die Deutsche Nationalbibliothek verzeichnet diese Publikation in der Deutschen Nationalbibliografie; detaillierte bibliografische Angaben sind im Internet über <http://dnb.d-nb.de> abrufbar.

Diese Arbeit hat der Fakultät für Elektrotechnik und Informationstechnik der Technischen Universität Ilmenau als Dissertation vorgelegen.
Tag der Einreichung: 7. April 2011
1. Gutachter: Univ.-Prof. Dr. rer. nat. habil. Andreas Schober
(Technische Universität Ilmenau)
2. Gutachter: Univ.-Prof. Dr.-Ing. habil. Hannes Töpfer
(Technische Universität Ilmenau)
3. Gutachter: PD Dr.-Ing. habil. Volker Cimalla
(Fraunhofer-Institut für Angewandte Festkörperphysik Freiburg)
Tag der Verteidigung: 22. Juni 2012

Technische Universität Ilmenau/Universitätsbibliothek

Universitätsverlag Ilmenau

Postfach 10 05 65
98684 Ilmenau
www.tu-ilmenau.de/universitaetsverlag

Herstellung und Auslieferung

Verlagshaus Monsenstein und Vannerdat OHG
Am Hawerkamp 31
48155 Münster
www.mv-verlag.de

ISBN 978-3-86360-055-6 (Druckausgabe)
URN [urn:nbn:de:gbv:ilm1-2012000298](http://nbn:de:gbv:ilm1-2012000298)

Titelfoto: photocase.com | AlexFlint

Contents

Symbols and Abbreviations	XXIII
1 Introduction	1
2 Theoretic Principles	7
2.1 Electrochemical Fundamentals	7
2.2 pH-Value and Buffer Solutions	17
3 Methods for the Measurement of pH-Value	25
3.1 Historical Development	25
3.2 Glass Electrode	30
3.3 Si-based ISFETs	35
3.3.1 Theory	36
3.3.2 Properties	42
3.3.3 Instrumentation	46
3.3.4 Applications	48
3.4 Reference Electrode	50
3.4.1 Conventional Design	50
3.4.2 Miniaturisation	54
4 AlGaN-based Fluid Sensors	59
4.1 The III-Nitride Material System	59

4.2	AlGa _N /Ga _N -Heterostructure and the Two-Dimensional Electron Gas	64
4.3	Defect-related Phenomena	66
4.3.1	Persistent Photoconductivity	68
4.3.2	Current Collapse	72
4.3.3	Defect etching	75
4.4	Applications as Fluid Sensors	77
4.4.1	Historical Background	78
4.4.2	Sensors for pH-Value	80
4.4.3	Sensors for Anions	83
4.4.4	Functionalised Sensors and Biosensors	90
4.5	Theoretical Description	91
5	Materials and Methods	95
5.1	Fabrication of AlGa _N -based pH-Sensors	95
5.2	Measurement Setups and Instrumentation for Impedance Analysis	100
5.2.1	Wafer-level Characterisation	102
5.2.2	Electrochemical Characterisation	102
5.2.3	Employed Buffer, Sample and Cleaning Solutions	106
5.3	Characterisation Methods	110
5.3.1	Impedance Spectroscopy	112
5.3.2	Impedance-Voltage Profiling	115
6	Impedance Characterisation of the AlGa_N/Ga_N Heterostructure and Derivation of Equivalent Circuits	121
6.1	Wafer level characterisation of the MIHS	122
6.1.1	Impedance Spectroscopy	122
6.1.2	Impedance-Voltage Profiling	139

6.1.3	Scanning Electron Microscopy	146
6.1.4	Summary	149
6.2	Chip level characterisation of the EIHS	152
6.2.1	Electrochemical Impedance Spectroscopy	153
6.2.2	Electrochemical Impedance-Voltage Profiling	161
6.2.3	Summary	164
7	Drift Behaviour and Sensitivity of the AlGaIn-EIHS	167
7.1	Characterisation of Drift Phenomena	167
7.1.1	Surface Contamination Drift	172
7.1.2	Light-induced Drift and Persistent Photoconduc- tivity	178
7.1.3	Electrochemical Corrosion	187
7.1.4	Summary	192
7.2	Sensitivity	193
7.2.1	High pH-Sensitivity	194
7.2.2	Diminished pH-Sensitivity	203
7.2.3	Summary	215
7.3	Cross-sensitivity to Ionic Strength	216
8	Optimisation of Ambient Conditions and Heterostructure Lay- out	227
8.1	Continuous Violet Illumination	228
8.1.1	Photoelectrochemical Corrosion with Violet Light	228
8.1.2	Characterisation of Light- and pH-Sensitivity of the Corroded Sensor in the Dark	241
8.1.3	Summary	249
8.2	Continuous Red Illumination	250
8.2.1	Photoelectrochemical Corrosion with Red Light .	251

8.2.2	Characterisation of Light- and pH-Sensitivity . . .	263
8.2.3	Summary	270
8.3	Increased cap layer thickness	271
8.3.1	Photoelectrochemical Corrosion	272
8.3.2	Characterisation of pH-Sensitivity	273
8.3.3	Summary	277
8.4	Reduced Al-content in the Barrier Layer	280
8.4.1	Photoelectrochemical Corrosion	280
8.4.2	Characterisation of pH-Sensitivity	283
8.4.3	Summary	285
9	Measurement of pH-Value and Development of a Micro-Blood- pH Analyser	287
9.1	Foetal Blood Sampling	288
9.2	Description of the Developed Micro-Blood-pH Analyser	290
9.3	Measurement of pH-Value with EIHS	299
9.4	Measurement of pH-Value with ISFETs	306
9.5	Summary	312
10	Conclusions and Outlook	315
	References	323